

## FLOW<sup>EVO</sup>

Infrared gas sensor CO<sub>2</sub> // Carbon dioxide // 30 Vol.-%  
smartGAS item number: F3-212307-05000



- Pre calibrated
- Compact design
- 3/5 mm gas line connectors
- 3.3 - 6 V DC supply voltage
- Modbus ASCII or RTU
- Status indication by LED
- Low drift

Non Dispersive Infrared (NDIR) gas sensor for process control and gas analysing using dual wavelength technology. Designed for process control, lab analysing and environmental monitoring in a wide range of gas measurement systems.

The FLOW<sup>EVO</sup> CO<sub>2</sub> sensor can easily be integrated into OEM systems, where long term stability, repeatability and reliable performance are required. Based on robust and precise NDIR technology our CO<sub>2</sub> sensors offer enduring solutions in the area of controlled combustion and process control. Furthermore, they can be used in environmental analysis and various other fields of scientific research where low signal drift and high selectivity are crucial for exact measurements and subsequent processing.

Modbus ASCII or RTU data communication offer a variety of options to connect the FLOW<sup>EVO</sup> sensor to a controller.

### APPLICATION EXAMPLE

GAS ANALYSING  
PROCESS CONTROL  
ENVIRONMENTAL MONITORING  
RESEARCH

**FLOW** <sup>EVO</sup>

Infrared gas sensor CO<sub>2</sub> // Carbon dioxide // 30 Vol.-%  
smartGAS item number: F3-212307-05000

General features	
Measurement principle:	Non Dispersive Infra-Red (NDIR), dual wavelength
Measurement range:	0..30 Vol.-% Full Scale (FS)
Gas supply:	by flow (nearly atmospheric pressure)
Flow rate:	0.1 .. 1.0 l / min
Dimensions:	73 mm x 30 mm x 36 mm (L x W x H)
Warm-up time:	< 2 minutes (start up time) < 30 minutes (full specification)
Measuring response*	
Response time (t <sub>90</sub> ):	Appr. 12 s @ 0.7 l / min
Digital resolution (@ zero):	0.01 Vol.-%
Detection limit (3 σ):	≤ 0.12 Vol.-%
Repeatability:	≤ ± 0.15 Vol.-%
Linearity error (straight line deviation):	≤ ± 0.25 Vol.-%
Long term stability (span):	≤ ± 0.6 Vol.-% over 1000 h period
Long term stability (zero):	≤ ± 0.4 Vol.-% over 1000 h period
Influence of T, P, flow rate, other*	
Temp. dependence (zero):	≤ ± 0.05 Vol.-% per °C
Temp. dependence (span):	≤ ± 0.08 Vol.-% per °C
Pressure dependence:	+ 0.156 % of measurement value / hPa
Flow rate dependence:	≤ ± 0.05 Vol.-% per 0.1 l / min
Cross sensitivity (zero) other gases:	consult factory
Electrical inputs and outputs	
Supply voltage:	3.3 V .. 6.0 V DC
Supply current (peak):	< 400 mA @ 3.3 V, < 240 mA @ 5.0 V
Inrush current:	< 600 mA
Average power consumption:	< 800 mW
Digital output signal:	Modbus ASCII / RTU via UART, autobaud, autoframe
Calibration:	zero and span by SW
Climatic conditions	
Operating temperature:	0 .. + 50 °C
Storage temperature:	-20 .. + 60 °C
Air pressure:	800 .. 1150 hPa
Ambient humidity:	0 .. 95 % relative humidity (not condensing)
* Typical values related to 1013 hPa, Ta=22 °C, flow = 0.7 l / min for dry (not condensing) and clean sample gas. Stated values exclude calibration gas tolerance.	

All rights reserved. Any logos and/or product names are trademarks of smartGAS. The reproduction, transfer, distribution or storage of information contained in this brochure in any form without the prior written consent of smartGAS is strictly prohibited. All specifications – technical included – are subject to change without notice. Depending on the application, the target gas and the measurement range the technical data may differ. No liability is accepted for any consequential losses, injury or damage resulting from the use of this document or from any omissions or errors herein. The data is given for guidance only. It does not constitute a specification or an offer for sale.

For more information, please visit [www.smartgas.eu](http://www.smartgas.eu) or contact us at [sales@smartgas.eu](mailto:sales@smartgas.eu)

Please consult smartGAS sales for parts specified with other temperature and measurement ranges.

At first initiation and depending on application and ambient conditions recalibration is recommended. Recurring cycles of recalibration are recommended.